

6.3.1. ต้องจัดทำแผนบำรุงรักษาเครื่องมือทางการแพทย์ตามคำแนะนำของผู้ผลิตเครื่องมือทางการแพทย์ หรือตามมาตรฐานของวิชาชีพที่เกี่ยวข้อง หรือตามประกาศของกรมสนับสนุนบริการสุขภาพ

ME CODE	Assigned to	Equipment	Manufacturer	Model	Serial Number	Department	Ri	Category	Decal	Status	Closed O	Comments
AMG-DEV-01	PREVENTIVE MAINTENANCE	AMALGAMATOR	3M	ESPE CAPREX	940036024219	DEV	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
AMG-DEV-02	PREVENTIVE MAINTENANCE	AMALGAMATOR	SDV	ULTRAMAT 2	101897	DEV	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
AST-ER-01(N)	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	YUWELL	7E-A	H2009070871	ER	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	3-Oct-25	-
SUD-IPD1-01	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	GASTRIC DRAINAGE	ONS	-	IPD1	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	1-Oct-25	-
VAC-LR-01	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	ATMOS	S 351 NATAL	1007549	LR	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	2-Oct-25	100,200,300 mbar
VAC-LR-02	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	ATMOS	ATMOPORTE 350	205532 02/2001-05	LR	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	2-Oct-25	0.1,0.5,0.9 bar
SUM-REF1-01	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	YUWELL	7E-A	H2009070895	REF1	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	3-Oct-25	-
SUM-REF3-01	PREVENTIVE MAINTENANCE AND CALIBRATION	ASPIRATOR	WEINMANN	ACCLIVAC LITE	-	REF3	LOW	Physical Therapy and Treatment Equipment	Approved	PASS	3-Oct-25	-
PRB-PT-01	PREVENTIVE MAINTENANCE AND CALIBRATION	BATH, PARAFFIN, PHYSICAL THERAPY	JMS	6.7D	14-014	PT	MEDIUM	Physical Therapy and Treatment Equipment	Approved	PASS	3-Oct-25	60
CEN-LAB-05	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	BOECO	C-28A	0001272-03	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	1000/2000/3000
SER-LAB-01	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	CLAY ADAMS	SERO-FLUGE2001	3680059	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	2940
CEN-LAB-02	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	BOECO	S-8	0000946	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	4000 3min
CEN-LAB-03	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	DIG-SYSTEM	DSC-158T	07091124	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	1x.6 นาทีต่อ
HEM-LAB-01	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	BOECO	H-240	0002449-02	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	12000 5min
CEN-LAB-06	PREVENTIVE MAINTENANCE AND CALIBRATION	CENTRIFUGE	BOECO	HC-240	001923-05	LAB	MEDIUM	Analytical Laboratory / Pharmacy	Approved	PASS	3-Oct-25	12000 5min
DEN-CUL-01	PREVENTIVE MAINTENANCE	Curing light	BA INTERNATIONAL	BF196	18D00076	DEN	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
DEN-CUL-02	PREVENTIVE MAINTENANCE	Curing light	BA INTERNATIONAL	BJ2FU	18D00079	DEN	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
DEN-CUL-03	PREVENTIVE MAINTENANCE	Curing light	DEMI	BGUZ6	752037468	DEN	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
DEN-CUL-04	PREVENTIVE MAINTENANCE	Curing light	3M	DEEPCURE-L	9321250609	DEN	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
DEN-CUL-05	PREVENTIVE MAINTENANCE	Curing light	3M	S10	939123-003600	DEN	LOW	Patient Related and Other	Approved	PASS	1-Oct-25	PM
AED-ER-01	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR, EXTERNAL, AUTOMATED (AED)	ZOLL	AEDPLUS	X16LR7585	ER	HIGH	Life Support	Approved	PASS	2-Oct-25	PAD EXPIED
AED-ER-02	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR, EXTERNAL, AUTOMATED (AED)	ZOLL	AEDPLUS	X21G385429	ER	HIGH	Life Support	Approved	PASS	2-Oct-25	PAD EXPIED
DEF-ER-03	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR/PACEMAKER, EXTERNAL	PHILIPS	EFFICIA DPM100	CN32629276	ER	HIGH	Life Support	Approved	PASS	1-Oct-25	-
DEF-IPD1-01	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR/PACEMAKER, EXTERNAL	PHILIPS	EFFICIA DPM100	CN32667157	IPD1	HIGH	Life Support	Approved	PASS	1-Oct-25	-
DEF-CHE-01	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR/PACEMAKER, EXTERNAL	PHILIPS	EFFICIA DPM100	CN32629287	IPD1	HIGH	Life Support	Approved	PASS	1-Oct-25	-
DEF-REF3-01	PREVENTIVE MAINTENANCE AND CALIBRATION	DEFIBRILLATOR/PACEMAKER, EXTERNAL	WEINMANN	MEDUCORE STANDARD	6148	REF3	HIGH	Life Support	Approved	PASS	3-Oct-25	ส่วน SPO2 ไม่สามารถ

แผนการสอบเทียบ และบำรุงรักษาเครื่องมือแพทย์ โรงพยาบาลสันกำแพง ปี 2569

ID CODE	ชื่อเครื่องมือ	หมายเลขครุภัณฑ์	ยี่ห้อ	รุ่น/แบบ (Model)	Serial Number	หน่วยงาน	Status	การดูแลรักษา	แผน
INP-CHE-30	INFUSION PUMP	6515-025-1001/41	TERUMO	TE-135	1606000182	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
INP-CHE-18	INFUSION PUMP	6515-025-1001/26	B BRAUN	INFUSOMAT P	121181	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
INP-CHE-09	INFUSION PUMP		B BRAUN	INFUSOMAT SPACE P	388794	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
INP-CHE-11	INFUSION PUMP		B BRAUN	INFUSOMAT SPACE P	386755	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
INP-CHE-07	INFUSION PUMP		B BRAUN	INFUSOMAT SPACE P	388786	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
INP-CHE-21	INFUSION PUMP		SK	SK-600II	SK9100356	CHE	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
DEN-CUL-01	Curing light	6520-009-0001/8	BA INTERNATIONAL	BF4Y6	18D00076	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
DEN-CUL-02	Curing light	6520-009-0001/9	BA INTERNATIONAL	BJ2FU	18D00079	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
DEN-CUL-03	Curing light	6520-009-0001/5	DEMI	BGUZ6	752037468	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
DEN-CUL-04	Curing light	6520-009-0001-1/67	3M	DEEPCURE-L	9321250609	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
DEN-CUL-05	Curing light	6520-009-0001/6	3M	S10	939123-003600	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
PUO-DEN-01	OXIMETER, PULSE		YUEWLL	-	2108015488	DEN	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...
WEA-DEN-01	SCALE, PATIENT, FLOOR	6530-008-0711/10	ZEPPER	ZT-160N	73988	DEN	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
ESU-ER-01	ELECTROSURGICAL UNIT	6515-035-0203/2(58)	LED	SURTRON 120	363019918	ER	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
LAR-ER-01(N)	LARYNGOSCOPE		-	-	-	ER	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
WEA-ER-01	SCALE, PATIENT, FLOOR		HOSPITAL & HOMECARE	-	-	ER	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
WEA-HW-01(N)	SCALE, PATIENT, FLOOR	6530-008-0711/8	ZEPPER	160 KG	-	HOME WARD	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
WEA-IPD1-01	SCALE, PATIENT, FLOOR		NAGATA	BW-1410H	F124064	IPD1	ใช้งานได้	PM+CAL	PM และ Calibration ทุก 1 ปี
WEB-IPD2-01	SCALE, INFANT	6530-008-0811/9	ZEPPER	EB-20	650101185	IPD2	ใช้งานได้	PM+CAL	PM ทุก 6 เดือน / Calibration ...